2009-03 Oxygen duplicates - page 1

Precision statement for replicate samples drawn from a single Niskin bottle:

The pooled standard deviation for Oxygen:Dissolved:Bottle for the range 0.192 to 6.962 ml/l was 0.012, k = 7 (1 outlier removed) where k is the number of pairs of duplicates.

Precision calculation for duplicate samples:

Precision was determined by analyzing replicate samples drawn from one Niskin. Pooled standard deviation of pairs of samples (Sp) was calucalated by:

 $Sp = SQRT\{sum (d^*d)/2k\}$

where k is the number of pairs and d is the difference between pairs.

Determination of outliers

Outliers are discarded on the basis of Chauvenet's criteria. The statistic is calculated by the difference between the outlier and the mean, divided by the stdev.

If this absolute value is greater than the critical value of the Chauvenet criterion for the given n, the datapoint can be discarded.

The Chauvenet Statistic assumes a normal distribution.

Duplicate analysis for samples drawn from a single Niskin bottle:

| Event Number | Sample Number | Station | Pressure dbar | Oxy:Dis 1 ml/l | Oxy:Dis 2 ml/l | Rejected yes / no |
|-----------------|------------------|---------|------------------|-------------------|-------------------|----------------------|
| 2 | 5 | Si03 | 124.5 | 0.384 | 0.199 | yes |
| 4 | 19 | P2 | 24.5 | 5.792 | 5.794 | |
| 9 | 62 | P4 | 799.4 | 0.192 | 0.216 | |
| 18 | 138 | P8 | 100.4 | 6.214 | 6.236 | |
| 27 | 212 | P12 | 74.5 | 6.714 | 6.698 | |
| 35 | 292 | P16 | 123.0 | 5.955 | 5.945 | |
| 38 | 368 | P20 | 24.1 | 6.962 | 6.937 | |
| 40 | 412 | P26 | 125.1 | 5.948 | 5.949 | |

2009-03 Oxygen duplicates - page 2

Precision statement for samples drawn from duplicate Niskin bottles closed at same pressure:

The pooled standard deviation for Oxygen:Dissolved:Bottle for the range 0.439 to 6.793 ml/l was 0.015, k = 4 (1 outlier removed) where k is the number of pairs of duplicates.

Duplicate Niskins at the same pressure

| Event Number | Sample Number | Station | Nominal Pressure dbar | Oxy:Dis 1 ml/l | Oxy:Dis 2 ml/l | Rejected yes / no | Comment |
|-----------------|------------------|---------|-----------------------------|-------------------|-------------------|----------------------|---------|
| 4 | 25 / 26 | P2 | 5 | 6.793 | 6.77 | | |
| 9 | 59 / 60 | P4 | 1250 | 0.449 | 0.439 | | _ |
| 27 | 196 / 197 | P12 | 2500 | 1.126 | 1.821 | yes | |
| 35 | 278 / 279 | P16 | 2500 | 1.882 | 1.901 | | |
| 38 | 350 / 351 | P20 | 2500 | 1.958 | 1.929 | | |

Note: even though the precision statement for samples drawn from duplicate Niskin bottles (not true duplicates) is calculated using the same formula as the precision statement for duplicates samples drawn from one single Niskin, this process is mainly used to identify problem samples and is not being used as a measure of analysis precision.