

2008-27 Total Alkalinity

Duplicate analysis for samples drawn from a single Niskin bottle:

Event Number	Sample Number	Station	Pressure dbar	ALK 1 µmol/kg	ALK 2 µmol/kg	ALK 3 µmol/kg	Difference µmol/kg	Absolute Difference	Standard Deviation	Comments
11	28	P4	1248.7	2395.20	2394.88		0.32	0.32		
38	193	P12	2999.8	2430.21	2429.19		1.01			Poor calibration
53	310	P16	3001.5	2430.31	2430.93		-0.62	0.62		
63	364	P20	3001.4	2429.01	2428.42		0.59	0.59		
72	429	P26	3500.2	2425.70	2424.53		1.17	1.17		
91	553	P25	1999.7	2420.85	2421.20		-0.35	0.35		
91	554	P25	2000.3	2421.20	2420.72		0.48	0.48		
91	555	P25	2000.2	2420.42	2421.26		-0.84	0.84		
91	556	P25	1999.6	2419.20	2419.02		0.18	0.18		
91	557	P25	1998.6	2420.36	2420.81		-0.45	0.45		
average variability:							0.15	0.56		
standard deviation of variability:							0.69	0.30		

Duplicate Niskins at the same pressure

Event Number	Sample Number	Station	Nominal Pressure dbar	ALK A µmol/kg	ALK B µmol/kg	ALK C µmol/kg	ALK D µmol/kg	ALK E µmol/kg	Difference µmol/kg	Absolute Difference	Standard Deviation	Comments
11	27 / 28	P4	1250	2394.05	2395.04				0.99	0.99		
38	192 / 193	P12	3000	2428.58	2430.21				1.62			Poor calibration
53	310 / 311	P16	3000	2430.62	2430.61				-0.01	0.01		
63	364 / 365	P20	3000	2428.71	2428.49				-0.23	0.23		
91	553 - 557	P25	2000	2421.02	2420.96	2420.84	2419.11	2420.59		0.80		
average variability:									0.59		0.51	
standard deviation of variability:									0.87		0.46	

Precision estimate: 0.56 µmol/kg