1.99958

Water Properties Group Institute of Ocean Sciences, Ocean Sciences Division

AUTOSAL ANALYSIS LOGSHEET

Cruise ID: 2012-09	Cast/Station Name:	Analysis Date (dd | mmm | yyyy): 5 | 12 | 12
Autosal Model: 8400B	Serial Number: 68572	
Initial Standby Value: 24+5985	Ratio After Standardization: 24+5975	

Cell Standaridization (Daily)	Standardization Values:	Standard Info.	
Conductivity must be ± 0.00001 compared to Standard	Before:	K_{15} VALUE: 0.99979	Bath Temperature: 24 °C
	After:	Batch #: P 153	Sample Temperature: 22 °C
	Notes:	Batch Date: 8 MAR/14	Room Temperature: 22 °C
			Analyst: H.M.

Sample	Station Name	Depth/Nisk#	Salinity 1	Salinity 2	Salinity 3	Comments
08/12	TSG-1		1.87815	1.87816		
09/12	TSG-2		1.88275	1.88279		
09/12	TSG-3		1.87838	1.87838		
09/12	TSG-4		1.82029	1.82030		
09/12	TSG-5		1.64789	1.64789		
09/12	TSG-6		1.78977	1.78979		
09/12	TSG-7		1.82221	1.82222		
09/12	TSG-8		1.84066	1.84066		
09/12	TSG-9		1.87803	1.87804		
09/12	TSG-10		1.70142	1.70144		
21/12	TSG-S01		1.56223	1.56221		
21/12	TSG-S02		1.61146	1.61146		
21/12	TSG-S03		1.58203	1.58204		
21/12	TSG-S04		1.59090	1.59092		
						Final Standby Value:

Notes: