

~~Sam 3, 9600, 8, None~~

Sea term cptobox / SBE Set to 9600 B
7 Bits
Even parity

Sea Sure

Connection in Sea Sure Set
to 9600, 7, 1, E

ground wire attached to
Sea water manifold.

Start file @ 2105 UTC
Aug 30 UTC

2340 UTC Sept 7 2015 2015-21 LaPerouse
- started file 002

1227 UTC on Sept the 8th.

The Wetlabs fluorometer was
cleaned, very little fouling was present
Scott Rose

1858 8 Sept UTC.

Noticed that the flow reads 1.6.
Get it back down to 1.0

MR

1738 14 Sept 2015 UTC 2015-21 LaP
-stopped logging file 002
