

AUTOSAL ANALYSIS LOGSHEET

Cruise ID: <u>2018-05</u>	Cast/Station Name: _____	Analysis Date (dd mmm yyyy): <u>15</u> <u>3</u> <u>18</u>
Autosal Model: <u>8400 B</u>	Serial Number: <u>68572</u>	
Initial Standby Value: <u>24+6090</u>	Ratio After Standardization: <u>24+6098</u>	
Cell Standardization (Daily)	Standardization Values:	Standard Info.
Conductivity must be ± 0.00001 compared to Standard 	Before: _____ After: _____ Notes: <u>6.50</u>	K ₁₅ VALUE: 0.999 <u>83</u> Batch #: <u>P160</u> Batch Date: <u>20/July/19</u>
		Bath Temperature: <u>24</u> °C Sample Temperature: <u>22</u> °C Room Temperature: <u>22</u> °C Analyst: <u>H.A.</u>

Sample	Station Name	Depth/Nisk#	Salinity 1	Salinity 2	Salinity 3	Comments
2	STA 59		1.75914	1.75909	1.75909	
11	59		1.76906	1.76906		
25	56		1.74827	1.74827		
39	46		1.74841	1.74842		
51	42		1.79877	1.79877		
74	39		1.76160	1.76158		
67	39		1.80513	1.8517		
81	39		1.48163	1.48167		
87	27		1.80129	1.80134		
104	02		1.79348	1.79348		
122	03		1.80265	1.80266		
138	BS17		1.69617	1.69618		
146	6		1.78668	1.78668		
159	9		1.78553	1.78555		
171	12		1.79016	1.79016		
188	14		1.78714	1.78714		
206	16		1.78032	1.78026	1.78025	
220	Van 1		1.74787	1.74788		
231	Van 4		1.70489	1.70490		
239	Van 5		1.66357	1.66358		
246	Van 6		1.55716	1.55715		
261	IND 8		1.57707	1.57708		
254	Ind 7		1.57905	1.57905		Error communicating with salinometer
271	Ind 4		1.59232	1.59234		"do you want to continue"
Surface	Vac 1		1.29410	1.29412		Final Standby Value:

Notes: _____

AUTOSAL ANALYSIS LOGSHEET

Cruise ID: <u>2018-05</u>		Cast/Station Name: _____		Analysis Date (dd mmm yyyy): <u>15 03 2018</u>	
Autosal Model: <u>8400 B</u>		Serial Number: <u>68572</u>			
Initial Standby Value: <u>24 + 6098</u>		Ratio After Standardization: <u>24 + 6098</u>			
Cell Standardization (Daily)	Standardization Values:	Standard Info.		Bath Temperature: <u>24</u> °C	
Conductivity must be ± 0.00001 compared to Standard 	Before: _____	K ₁₅ VALUE: <u>0.999</u>		Sample Temperature: <u>22</u> °C	
	After: _____	Batch #: _____		Room Temperature: <u>22</u> °C	
	Notes: _____	Batch Date: _____		Analyst: <u>PJB</u>	

[illegible]

Notes: _____