

(16)

Peltier traps - please add back  
plate so fingers can't get in the  
Fans.

(17)

S.Z.

1 Feb 2019 - Mostly fixed it all up!

Filename: has 1354 as time

Testing - Reset sample time to 20sec  
(from 60sec) so results  
would be FASTER

There is still some leak in system  
letting room air in as shown by  
blowing on connections.

Base value  $\approx$  439 (XC) 480

Running STDs: -1; 370ish, 708

Blow around boxes ports (outsides)  
gives top of equilibrator  
connections in  
"FROM SHEQ" +  
"TO SHEQ"

can raise value to  $\approx$  518

Hard to figure out which is  
culprit but the area around box  
seems OK.

So 25/500  $\approx$  5% error ...

Stopped file and reset sample interval  
to 60sec and restarted program  
@ 1446

4 July (5 July 0335 UTC)

Starting up pCO<sub>2</sub> system:

STDS set at 10PSI  
O<sub>2</sub>, 384, 1138 ppm

Water coming at 0.5 ~~lpm~~  $\Delta$  0.5  
\* Flow rate not yet measured

Settings

LI 840 Com Port 3

VICI Valve Com Port 5

NI-DAQ

Serial output cable not enabled

Sample Int 2

# STDS 3

1st Std Time 120 sec

Post std delay 1 min

Std duration 90 sec

Min time btw stds 240 min

No TSG

12V I/O 1 is on

Saving File to 2019\2019-78-Dat1\F

Readings:

See Atm (too high)

STDS are O<sub>2</sub>, 2380, 2100 but not too stable

Sample is 400 to 500 ppm

Flow ball is pegged at top ... at least during sample measurements.

Temp is -3.8C, should be ~15C

Pressure is 75.4 kPa (units? Book's example shows -20 kPa)

LICOR DATA: xc 428

CO<sub>2</sub> Abs 0.09

Analo I/O

H<sub>2</sub>O 12.42

0.93 12V I/O  
FLOW meter

H<sub>2</sub>O abs 0.08

Cell Temp 51.90C

@ 0814

Cell Pres 102.60

July 4th

Input Voltage 24.07

Airp time completely off. Need to reset. 0815 4th July = 0441 5th July OR  
pCO<sub>2</sub> correct

- ① Shut down software + pCO<sub>2</sub> box
- ② UTC 0507 July 5th Res Set time on ~~sys~~ computer clock
- ③ The water level in the SHEQ had been slowly rising. When system powered off the water level went back down to the blue-taped lines.
- ④ Unplugged and replugged temperature connector to box.
- ⑤ Before turning off see that breathing into SHEQ holes gives huge spike. <sup>11000 ppm</sup> What are these holes in the side, just above the blue tape? Maybe they shld be plugged?

So water level in SHEQ was rising when sample pump was on.

I could try opening the make up air?

Turn on again: File 2014-78-data2 @ 0519

- Water level jumps back up to where it was
- Temperature still is -3.8
- Atm 550-600
- STD 1=0, 380, 1119
- STDS are much more steady this time around.
- (Flow bill passed even during <sup>STDS</sup> gas sampling + when turned off.)

Covering side SHEQ barrel holes the pCO2 drops from 500 to ~475

Opening the make up air to the lab gives ~505 pCO2 ppm. Water level does not drop so its not a pressure thing? Breathing into it I get a good spike

Com 3 Licor 5 Valve NI DAQ AI channels Dev/al 207

Serial output 4

! 0540 - covered the SHEQ holes w/ tape. now @ ~482 ppm -> 475

But breathy all our connectors still gurgled it reading up to 550 so still some leaks...

0930 - check in w/ system. Lab empty, reading ~410 ppm

taped Breath test... ① connector btw SHEQ and Peltier gives a +15

taped ② connection to Peltier give +40

③ exit from water trap one-way filter gives +20 But other connections seem ok or only minor

0955 - then tried to drain Peltier trap - under suction so water would not drain, just suck air in. I guess pump needs to be off to drain sample

Hmm taped these connectors but they are still leaking

1015 - stop for now

(22)

5 July 2033 - Fiddly again  
unplug To SHEQ

6 July 1316 local (GMT -7)

Sample + air line don't have  
restrictor valves on the  
small lines going b/w "T" and  
multipoint valve.

I stole the restrictor valve off  
STD 4 (unused) and added it to  
the sample line and now flow  
is no longer > 250 ml/min and  
has been reduced to 50 ml/min

This will reduce volume of makeup  
air and should help w/ the back-  
pressure in the SHEQ

Next Peltier A cools well but leaks  
at valves  
Peltier B isn't cooling well, but  
has nice connectors...

↑  
so the nice connectors on Peltier B  
removed and swapped w/ A.

(23)

Peltier A back in place w/ sample  
Flow

The water block filters removed.

- IF they get wet will pump blow up due to  
~~no~~ no flow?
- Are they making it harder for flow to  
travel creating pressure @ SHEQ?

Restart software @ 2210 UTC July 6

STDs 1, 3 & 1120  
Water is 385 (!) much lower than before  
Atm is 500 so clearly in a  
contaminated spot

More changes:

Water level in SHEQ was rising so  
wondered if make up air was  
not actually available from the tube  
it was in.

Changed plumbing to bring make-up  
air from atmosphere exhaust T,  
though since atm line is contaminated  
sample value is still too high.

Opened "To SHEQ" restrictor flow  
to all the way open + then readjusted  
the small restrictor valve to multipoint valve  
until flow was ~50 ml/min

(24)

So correct values @ 22:19: ~~19~~ 6 Jul

XC 380.54

CO<sub>2</sub>Ab 0.08

H<sub>2</sub>O 10.95

H<sub>2</sub>O<sub>obs</sub> 0.07

Cell T 51.89

Cell P 102.58

Input V 24.04V

Temp -3.7

Press 75.3

1 Anal. V<sub>o</sub> 0.96

2 " 0

3 " 0

Time Stamp (good) 22:21

Sample Flow @ meter  
50 to 70 ml/min

Make up air from  
Air exhaust T

Peltier A w/ no  
water block  
Filter

quick  
check  
connectors

~~Quick check for~~

SHEQ barrels  
still have  
tape over  
"Atm vents"

Water level in SHEQ  
is at top of blue tape mark for each  
barrel.

Flow is at 0.5 Gallons per min  
for each barrel based on flow gauges

Peltier B NOT set up on Atm line.

(25)

Breath test: In box OK  
outside box OK  
top of SHEQS OK  
Peltier A OK

Stable at 387 22:33 UTC

STD 1 Flow restrictor valve was on  
backwards so this was undone +  
changed direction so that

symbol matches others  
at the

7 July 03:28 UTC 03:35 UTC

Disconnected the atmosphere line  
and

Connected new line that runs out  
the window/davit ~~room~~ control room  
into very fresh air

CO<sub>2</sub> had been ~410 (400 to 420)

As make up air from atm line getting sucked in,  
the <sup>sample</sup>CO<sub>2</sub> is dropping. Levels out at 390 ppm.

How much STD?

0 : ~~950~~ 950 psi set @ 10psi

384 : ~~12,000~~ 1200 psi set @ 10psi

1138 : 1850 psi set @ 10psi

0345 Renn STDs! @ 0345

- ① Atm 405
- ② NO  $\phi$ ? NO FLOW? why was valve in back?
- ③ 384 <sup>cool!</sup> coming in at over 250 ml/min
- ④ 1122 " " " " "

check resistor values tomorrow...

Unit sample now coming in over 250ml  
So at some point valve was passed +  
Atm may have passed so stuck

still w/ atm line, dropping down to 381ppm  
0402 - down to ~~327~~ 377pp @ 0402  
but then needed to pull the atm line in

1509 Took tape of the SHEQ vents. ~~be~~  
These are ~~rather~~ "antisipon" parts to  
keep SHEQ from accidentally overflowing.  
Now that pressure is not being pulled  
in there is no longer so much suction

in SHEQ the vents won't be actively drawing  
in air. I hope.  
lab

1611 @ 1611 adjusted Flow restrictors on all  
the standards so now there is flow on  
0 STD again! and they are all at  
100 ml/min.

2017 Chef F gave permission to run atmospheric line  
through main lab + ctd lab doors. We then  
ran line out ctd lab window to Rosette area,  
up to deck above, Forward, then up to bridge  
level, just aft of bridge house - just outside  
the port door.

Plugged into box - although a run of ~50m  
the atmosphere exhaust feels strong.  
Atmospheric line is a) ~45m of black  
stiff tube, similar diameter to the BEVA-IV  
tubing. b) ~5m of BEVA-IV tubing and  
c) ends with 6" of flexible tubing +  
funnel cup.

Very few connections. Nice!

STDs had just run, so set it to  
run again and monitor flow.

No spikes w/ break test.

Value:	406	3	382	<del>1122</del>	sample @ 398
Flow:	7250	100	100	100	50
wh <sup>at</sup> :	AIR	0	①	②	sample

Peltier A is filling up fast!  
 At ~24 hr is 2/3 Full  
 20:35 Empty @ 2035  
 Had to disconnect "Peltier to line to box"  
 so there would be to empty, stirring  
 its just sucking in air bubbles through  
 the Peltier knob.  
 ↑ open  
 ie pump is pulling air through system  
 so need to disengage the pump  
 suction to empty the Peltier.  
 Breath test in front of <sup>untaped</sup> ~~SHEA~~ <sup>atm</sup> vents  
 has NO effect! Yay!

So what's left?  
 • Fix Peltier B (doesn't get cold)  
 • See if wiring is compromised for  
 pressure + temperature  
 though these #s do change ie 3.8 to 4.1  
 kPa 75.2 to 75.5

8 July 0930 - Must have had a leak in  
 the STD line. ~~It~~ is the tank  
 is now empty!

1008 emptied Peltier trap.  
 Had been breathing on system, seemed to  
 vary 411 → 420 ... but not sure. No changes

started sucking w/ line because  
 sample went from 390 to 420 after  
 stds run...

9 July 0417 - emptied Peltier trap.

1948 - after taking apart sample T  
 water was pulled in pump (but  
 came out open T so no water into  
 Lixer.  
 Sent some FW through pump, changed  
 tubing, ran pump more, still had  
 water in it so waited 'til morning.  
 Borrowed ET's "torque-star shaped wrench",  
 opened pump, wiped down + made sure  
 it was dry, reassembled and started up  
 @ 1948.

407 Serial OUTPUT was not enabled but  
~~AA~~ tried setting to COM4 and enabling.  
 326 364 not sure what this does?

Values:	CO <sub>2</sub>	A <sub>g</sub> H <sub>2</sub> O in line	Flow
sample	390	11	50
air	407	○	7250
STD1	—	—	○
STD2	364	7 ish	100
STD3	1119	10.8	

and I put the water trap filter back in line.

Note there is no Peltier or water trap on  
 air currently. Need to add

380 STD fault is @ 1800 PSI

10 July 2019

1600 UTC -

Breath on SHEQ + trap no change  
Breath on outside of box ~~36~~ 359 to 362  
so slight leak here.  
Breath on inside of box as well 356-364

Close back up @ 1609  
Pushed all connectors back together in to check tightness. Did not rewire any.

Peltier trap is 1/2 full emptied using first tubing off the "To Box" line.  
SPIKE to 2600 ppm (wow!)

also sample flow was down to 40 ml/min so this was increased using flow restrictor. Hmm, increased it to 50, but now after emptying the Peltier its up at 100 ml/min.

Open <sup>box</sup> again and adjust flow. Hmm not needed it was just a spike I guess.

1617 stopped fiddling.

ATM / STDs :

Atm 409

STD1 X

STD2 30 → 383

STD3 11??

} looks good!

10 July 2139 UTC

System turned off + seawater loop turned off as we approach Dutch Harbor.

11 July 2331 UTC

seawater loop turned on + pe

Check STDs - tanks were turned off when system shut down yesterday.

Today 384 - still holding 10psi + 1800psi

1138 - was 0, 0 ton on and now 10psi + 1800psi

\* So 1138 ppm line has leak somewhere

seawater @ 0.5 gpm each

Removed sample water trap/Filter (Vocoguard) - worried it creates too much suction...

Reboot tablet, power back licor + pumps on, Settings - change file

Sample flow @ 50 to 60 ml/min

Settings: Change File size to 240min so there will be a set of STD for each file.

Change STD duration to 100 sec. to help stabilize STD 2 (384ppm) -

Start @ 2352

Xc: 306  
CO2 Abs 0.07  
H2O 11.15  
H2O Ab 0.07  
Cell T 45.66  
Cell P 101.35 kPa  
Input V 23.95V  
T -5.0  
P 74.7 kPa

Analog I/O 0.90 <sup>12V</sup> / 101  
Time 42 seconds faster than GPS

	CO2	H2O	Flow
Sample:			
atm:	430	12.7	7250ml
STD 1	420	13.5	7250ml
2			100ml/min
3	1112	2.13	

Settings: STD duration back to 90sec  
Post-STD delay 1min  
Time btw STD 240min  
Equilib? display 1hr  
Licor }  
Data file duration 90min

I had changed those values starting the software but now they are back to what they were...

Licor still waking up? Re do STDs all in a bit...

12 July 0144 a couple hrs later...

CO2 32.7  
H2O 10.98  
Cell T 51.89  
Cell P 101.37  
Input V 24.10  
Analog I/O 0.86V  
T -4.9  
P 74.9 kPa  
sample flow at 70-80ml/min  
water flow at 0.5 gpm each

0455 - tried snooping for leak in 1138 ppm line but didn't see anything. Next exp: turned off tank again to see how long it takes for 10psi to drop

34

6607 UTC - OK line has gone to 0 psi but regulator side still ~~at~~ good. so leak in btw regulator + line...

Found leak in copper line connector. ~~Removed~~ Tried to clip off extra copper tubing going into connector but squeezed the tube shut. darn.

0730: 310 so cut of ~~a~~ cone bit w/ hacksaw and put on new washer + cone and wrenched it tightly shut.

Flow too high → 383 after rebond 1164 w/ 120 ml/min flow (creeping up)

0811 Snoop shows no more leak there. After STDS run, turned off regulator again to confirm no remaining leak. Only waited to 0811 - looks like pressure may have been falling slowly... rats.

1839 - Empty Pellet  
1915 - Run STDS & exit program  
Turn off ~~power~~ power / pump  
Empty Pellet  
Remove, trim tubing, for all sample-line connectors: though did not muck w/ multipoint valve or loose connecting

35

For 2019:

- SNOOP
- SPARE PUMP
- More air flow meters?
- More connectors for tubing (copper or plastic or PVC)

2058 - Restart software

All values high...

Sample CO<sub>2</sub> ~ 360

Air ~ 420 ⇒ we had fire drill &

STD 1 - X two doors closed on line

2 - Flow still good but may

3 - 1114 <sup>and rising</sup> have issue w/ leaks/breaks?

come back and run STDS in a little bit...

13 July 0049

Sample

Air	409	7250	12.5
1	-	?	6.8
2	384	?	1
3	1116	?	

• UNSTICK FLOWMETER  
• Δ to 120 sec STD  
Retake STDS:

409	7250	12.4
Ball stuck	∅	12.5
386	150	4.3 → 0.9
1122	125	0.5

Need ✓

13 July Hill

- peltier 1/2 full trying to empty it.
- pCO<sub>2</sub> in position 1 to sucking up air & water doesn't empty.

~17:30 so seawater loop flow completely stopped so shut off pCO<sub>2</sub> and gases

14 July 0144 UTC

- seawater pump fixed.
- Case A second gas leak in 1138 STD Found + fixed (other side of previous leaky connection)
- Gas turned on.
- Systems + plumbing turned back on while on station SLIP-1

	CO <sub>2</sub>	Flow	H <sub>2</sub> O	STARTING
Sample	275	50 ml/min	13	STDs
Atm	418	7250	13	
STD1	-	0		
2	373	250 ml/min	0	
3	1082	100 ml/min	1.3	

STD 2 needs flow restrictor adjusted still.

2027 Run STD<sub>A</sub> + empty Peltier, and restart STOP Acquisition

Sample at ~180 ppm !!

pos		CO <sub>2</sub>	Flow	H <sub>2</sub> O
2	Atm	407	100 ml/min	13.13
3	STD1	406	100 ml/min	13.20
4	STD2	378	"	
5	STD3	1114	100 ml	1.4

Sample Flow @ 100 ... seems like its stuck.

yes pump turned off and bill still @ 100 ml/min!

- Unstick Flow bill by blowing down exit hole.
- Restart pump + software
- empty Peltier Trap

Sample Flow @ 70 ml/min and value still @ 191 ppm

15 July 0647

Run STDs, unstick Flow meter:

	CO <sub>2</sub>	H <sub>2</sub> O	Flow
Atm	406	11.2	7250 ml/min
STD1	400	0	
2	382	0.4	120
3	1113	1	110

Power OFF, EMPTY PELTIER (only 1/4 full)

Restart power + software

Sample back to 212 ppm

16 July 2019  
0906 - Run STDS + drain peltier

Sample	190	and rising as we steam	50ml/min	10.5
AIR	406	7250		9.7
STD1	407	0		9.7
2	375 → 383	120		6.7 → 6.5
3	1111 → 1115	110		1.1 → 0.8

- Collection stopped
- Pump off
- Peltier emptied
- Pump on
- Restart program

Display:

cell T 51.87  
 P 100.59  
 In V 24.10  
 Temp -6.6 C  
 KPa 73.7  
 12V 1/0 0.78 (0.5 GPM each)

Gasses 1138 @ 10 psi e 1800 psi  
 384 @ 10 psi e 1150 psi

16 July 2120: STDS, Power off, empty peltier, restart power + software

17 July 1952:  
A flow rate from 0.45 to 0.50 and see drop in pCO2 from 246 → 228 ppm. Drop was immediate. ? why so sensitive?

20:05 - see that <sup>sample air</sup> flow was low ~ 30 ml/min. Changed this to 50 using the flow restrictor. The CO2 increased from 228 to 232

Peltier empty process: STDS, stop, empty, start

CO2	Flow	H2O	Sample
232	50ml	11	Sample
408	2250	10.1	Atm
409	Out/min	10	STD1
378 → 383	120	5.6 → 12	STD2 ← Value closer to actual as it dries out
1111 → 1117	100	93 → 0.8	STD3

STOP + RESTART  
Sample flow at 50ml/min and 0.5 GPM each  
STDS run and now sample @ 223 ppm

(40)

18 July

0843 - on SEC 7 station

stds run  
then test of flow rate...

Sample at 352 <sup>air</sup> <sup>water</sup> <sup>flow</sup>  
Flow at 50 ml/min, 0.5 gpm, 0.85V

increase to 0.6 gpm (osch), ~~1.04V~~  
hmm, sample is slowly dropping... 1.03V  
to 348 ppm

0903: And set back to .5 gpm, 0.86V

1702: Flow was @ 40 ml/min  
adjusted to 60 ml/min

1926: Run stds + empty Peltier trap

Sample 220ish  
Atm 402

STD2 374 6 → 381 4 → 382 2  
STD3 1110 1.0

1929: Result

Sample 210<sup>200</sup> 11.5 <sup>set</sup> <sup>stds</sup>  
Atm 405 12.5 <sup>120 sec</sup>

STD1 411 12.4  
2 382 3.6 → 384 2.5 384 1.0  
1016 1.1 1114 0.9 → 1119 1.75

1932 adjusted flow - was at 0.6 GPM?  
~~was~~ down to 0.5 GPM

19 July 2019

0825 - Flow 0.5 GPM

sample air flow 45 to 55 ml/min  
Brenth test in the box

raises value 249 to 252  
so not much.

0828

empty peltier trap - stds run less  
than an hour ago so no need to  
re-run before emptying trap.

increase flow to 55 ml/min

20 July 2019

0741 - Peltier emptying

	CO2	Flow	H2O	water
Sample	261	(40)	10	0.5 0.79V
Air	404	7250	8.35	
STD1	404	∅	8	
2	382 → 383	100	4.9 → 1.5	
3	1108 → 1117	100	1.0 → 0.6	

Gases @ 1200 for 384, 1800 for 1138  
both @ 10 psi

20 July 2105  
stopped program / pump / emptied petir

sample flow @ 50 ml/min

21 July 1048

Sample flow @ 30 ml/min

TST water flow @ 0.45 g per min  $V=0.66$

increased to 0.5 g per min (and  $V=0.83$ )

RUN STDS / POWER OFF / EMPTY PELTIER /  
Restart

ADJUST FLOW OF sample

	Flow ml/min	H <sub>2</sub> O #
Sample	~ 30	
ATM 408	> 250	7.5
STD 1		
2 385	110	5.9
3 118	110	0.5

checked STD before stopping

AFTER restart adjusted sample flow.  
Saw increased flow but then to  
bring back to 50 seems like dial  
restriction valve knob was almost back  
to where it was ... maybe just a small small  
block in flow? water drop that made its  
way past?

22 July 1141

H<sub>2</sub>O for sample is up at 19! ?  
not sure what happened ... Peltier trap is  
only 1/3 + 1/2 full ...

Flow is at 0

Run ATM + STDS ... empty trap.

Maybe from ATM line?  
water

Sample	VP <sub>ATM</sub>	Flow	H <sub>2</sub> O
Air	396	> 250	11
STD 1			
2	383	110	1.2
3	115	110	0.6

1242 - finished mucking w/ plumbing

- Flow of sample was 0  
but flow up to multiport could be felt.  
other sample gases running through Licox Five so  
seemed to be multiport valve
- unconnected + reconnected multiport and  
then flow was great

③ but when hooking air line  
from peltier trap back in (closed  
the loop) the flow dropped  
right down ... something blocking  
flow ...

- so examining lines found the  
plumbing from SHEQ to peltier was  
crusted inside w/ salt.  
Disconnected, put DI water  
through, tapped out water as  
best as possible & reconnected.

~~It~~  
This seems to have fixed it.  
Flow ~~is~~ set pt is back to where  
it was.  $H_2O$  value has  
dropped back down to 11.

back down to  
Sample value @ 317 and  
maybe still falling.

1250 Started new file. ~~Low~~  
of good data. ~~the std loss the~~  
~~last~~ Don't use the previous  
hour's data.

2001 - Last STDs and turned system  
off. Flush w/ fw and pack.

END OF CRUISE