Water Properties Group Institute of Ocean Sciences, Ocean Sciences Division

AUTOSAL ANALYSIS LOGSHEET

Cruise ID: 2022-021	Cast/Station Name: Many	Analysis Date (dd | mmm | yyyy): 10 | Nov | 2022
Autosal Model: 8400B	Serial Number: 73274	
Initial Standby Value: 24+ 5979	Ratio After Standardization:	

Cell Standaridization (Daily)	Standardization Values:	Standard Info.	
Conductivity must be ± 0.00001 compared to Standard	Before:	K_{15} VALUE: 0.999	Bath Temperature: 24 °C
	After:	Batch #:	Sample Temperature: ~22 °C
	Notes:	Batch Date:	Room Temperature: ~22 °C
			Analyst: KS

Sample	Station Name	Depth/Nisk#	Salinity 1	Salinity 2	Salinity 3	Comments
59	NC14	2	1.75791	1.75790		
95	PA7	5	1.73796	1.73795		
10A	NC7	2	1.75780	1.75778		
16	"	8	1.74419	1.74420		
92	PA7	2	1.75064	1.75063		
10B	NC7	2	1.75778	1.75777		
DWR			1.97867	1.97868		

59	95	10A	16	92	10B
SAL	SAL	SAL	SAL	SAL	SAL
NC14	PA7	NC7	NC7	PA7	NC7
2022-021	2022-021	2022-021	2022-021	2022-021	2022-021
250 m	10 m	200 m	10 m	Bot-10 m	200 m
bot 2	bot 5	bot 2	bot 8	bot 2	bot 2

Final Standby Value:

Notes: DWR 2022-008 P23 2005 m bot 1, 2