

AUTOSAL ANALYSIS LOGSHEET

Cruise ID: <u>2023-010</u>		Cast/Station Name: <u>many</u>		Analysis Date (dd mmm yyyy): <u>13 March 2023</u>	
Autosal Model: <u>8400B</u>		Serial Number: <u>73274</u>			
Initial Standby Value: <u>24 + 5979</u>		Ratio After Standardization: <u>24 + 5985 4</u>			
Cell Standardization (Daily)		Standardization Values:		Standard Info.	
Conductivity must be ± 0.00001 compared to Standard	Before: <u>4.79</u>		K ₁₅ VALUE: 0.999 <u>86</u>		Bath Temperature: <u>24</u> °C
	After: <u>4.875</u>		Batch #: <u>P165</u>		Sample Temperature: <u>~22.5</u> °C
	Notes: _____		Exy Batch Date: <u>15 April 2024</u>		Room Temperature: <u>22.4</u> °C
					Analyst: <u>KS</u>

[illegible]

Notes: _____