Water Properties Group Institute of Ocean Sciences, Ocean Sciences Division

AUTOSAL ANALYSIS LOGSHEET

Cruise ID: 2023-010_Sept	Cast/Station Name: Many	Analysis Date (dd | mmm | yyyy): 28 | Sept | 2023
Autosal Model: 8400B	Serial Number: 73274	
Initial Standby Value: 24+ 5980	Ratio After Standardization:	

Cell Standaridization (Daily)	Standardization Values:	Standard Info.	
Conductivity must be ± 0.00001 compared to Standard	Before:	K_{15} VALUE: 0.99987	Bath Temperature: 24 °C
	After:	Batch #: See pag 3 of F	Sample Temperature: ~22 °C
	Notes:	Batch Date: 2023-026	Room Temperature: 22.2 °C
			Analyst: KS

Sample	Station Name	Depth/Nisk#	Salinity 1	Salinity 2	Salinity 3	Comments
904A	Swale	1	1.90819	1.90818		
901B	Sarita	1	1.90698	1.90697		
904B	Swale	1	1.90810	1.90809		
905	Swale	2	1.81705	1.81706		
901A	Sarita	1	1.90697	1.90698		
902	Sarita	2	1.78213	1.78212		
						— END

904A	901B	904B	905	901A	902
Sal	Sal	Sal	Sal	Sal	Sal
Swale	Sarita	Swale	Swale	Sarita	Sarita
2023-010	2023-010	2023-010	2023-010	2023-010	2023-010
Bot-5 m	Bot-5 m	Bot-5 m	0 m	Bot-5 m	0 m
bot 1	bot 1	bot 1	bot 2	bot 1	bot 2

Final Standby Value:

Notes: