

AUTOSAL ANALYSIS LOGSHEET

Cruise ID: <u>2023-068</u>	Cast/Station Name: <u>Mary</u>	Analysis Date (dd mmm yyyy): <u>06 Sept. 2023</u>
Autosal Model: <u>8400B</u>	Serial Number: <u>73274</u>	
Initial Standby Value: <u>241.5979</u>	Ratio After Standardization: _____	
Cell Standardization (Daily)	Standardization Values:	Standard Info.
Conductivity must be ± 0.00001 compared to Standard	Before: _____ After: _____ Notes: _____	K ₁₅ VALUE: 0.999 <u>86</u> Batch #: <u>see page 1 of</u> Batch Date: <u>2023-088</u>
		Bath Temperature: <u>24</u> °C Sample Temperature: <u>22.2</u> °C Room Temperature: <u>22.4</u> °C Analyst: <u>KS</u>

Sample	Station Name	Depth/Nisk#	Salinity 1	Salinity 2	Salinity 3	Comments
101	CC36	1	1.78366	1.78365		
114A	"	14	1.75138	1.75137		
114B	"	14	1.75124	1.75122		
30	RP-6C	1	1.83817	1.83819		
29A	RP-5C	6	1.73069	1.73065		
1	RP-1C	1	1.83958	1.83956		
72A	R2	6	1.78356	1.78355		
9B	RP-2C	1	1.82846	1.82843		
61B	C26	1	1.83401	1.83405		
42B	C12	6	1.79338	1.79340		
49A	C1	1	1.82855	1.82859		
61A	C26	1	1.83401	1.83399		
9A	RP-2C	1	1.82858	1.82860		
36B	RP-6C	7	1.76593	1.76595		
37	C12	1	1.82024	1.82023		
8B	RP-1C	8	1.70500	1.70501		
14	RP-2C	6	1.71662	1.71663		
36A	RP-6C	7	1.76565	1.76565		
66	C26	6	1.78791	1.78790		
8A	RP-1C	8	1.70497	1.70496		
24	RP-5C	1	1.83111	1.83112		
67	R2	1	1.82616	1.82615		
72B	R2	6	1.78314	1.78311		
29B	RP-5C	6	1.73080	1.73077		
42A	C12	6	1.79340	1.79344		Final Standby Value:
54	C1	6	1.79024	1.79021		

Notes: <u>49B C1</u>	<u>1.82769</u>	<u>1.82768</u>
<u>DNR</u>	<u>1.97831</u>	<u>1.97829</u>
	<u>1.97831</u>	<u>1.97829</u>

Labels

06 September 2023

101

[illegible]

Notes: _____