

AUTOSAL ANALYSIS LOGSHEET

Cruise ID: <u>2023-091</u>		Cast/Station Name: _____		Analysis Date (dd mmm yyyy): <u>1 Nov 2023</u>	
Autosal Model: <u>8400B</u>		Serial Number: <u>73274</u>			
Initial Standby Value: <u>2445979</u>		Ratio After Standardization: _____			
Cell Standardization (Daily)	Standardization Values:	Standard Info.		Bath Temperature: <u>24</u> °C	
Conductivity must be ± 0.00001 compared to Standard 	Before: _____	K ₁₅ VALUE: <u>0.99987</u>		Sample Temperature: <u>22</u> °C	
	After: _____	Batch #: <u>P166</u>		Room Temperature: <u>22.2</u> °C	
	Notes: _____	Batch Date: <u>5 Apr 2025</u>		Analyst: _____	

[illegible]

Notes: using cal from 2023-009

Cruise ID: _____		Cast/Station Name: _____		Analysis Date (dd mmm yyyy): _____	
Portasal Serial Number: _____		Reference Calibration Values: _____		_____ _____ _____	
Initial Standby Value: _____		- _____ + _____ ± _____		Zero Value: _____	
Cell Standaridization (Daily)	Standardization Values:	Standard Info.		Bath Temperature: _____ °C	
Conductivity must be ± 0.00001 compared to Standard	Before: _____	K ₁₅ VALUE: 0.999 _____		Sample Temperature: _____ °C	
	After: _____	Batch #: _____		Room Temperature: _____ °C	
	Notes: _____	Batch Date: _____		Analyst: _____	

[illegible]

Notes: _____