

Cruise ID: <u>2024-009</u> <u>Apr</u>		Cast/Station Name: _____	Analysis Date (dd mmm yyyy): _____
Autosal Model: <u>8400B</u>		Serial Number: <u>73274</u>	<u>25</u> <u>Apr</u> <u>2024</u>
Initial Standby Value: <u>24.5986</u>		Ratio After Standardization: _____	
Cell Standardization (Daily)	Standardization Values:	Standard Info.	Bath Temperature: <u>24</u> °C
Conductivity must be ± 0.00001	Before: _____	K ₁₅ VALUE: 0.999 <u>87</u>	Sample Temperature: <u>23</u> °C
compared to Standard	After: _____	Batch #: <u>P166</u>	Room Temperature: <u>22.8</u> °C
	Notes: _____	Batch Date: <u>6 Apr · 2025</u>	Analyst: <u>CI</u>

[illegible]

Notes: used Std front 2024-043