

Cruise ID: <u>2024-005</u>		Cast/Station Name: <u>Barkley</u> <u>Feb</u>	Analysis Date (dd mmm yyyy): <u>20</u> <u>Feb</u> <u>2024</u>
Autosal Model: <u>8400 B</u>		Serial Number: <u>73274</u>	
Initial Standby Value: <u>24 + 5984</u>		Ratio After Standardization: <u>-0.000003</u>	
Cell Standardization (Daily)	Standardization Values:	Standard Info.	Bath Temperature: <u>24</u> °C
Conductivity must be ± 0.00001 compared to Standard	Before: _____	K ₁₅ VALUE: <u>0.999 86</u>	Sample Temperature: <u>22.0</u> °C
	After: _____	Batch #: <u>P1165</u>	Room Temperature: <u>21.0</u> °C
	Notes: _____	Batch Date: <u>15-Apr-2024</u>	Analyst: <u>CI</u>

Notes: standardization from yesterday 2024-002