

Cruise ID: <u>2024-005</u>		Cast/Station Name: <u>Mary</u>	Analysis Date (dd mmm yyyy):
Autosal Model: <u>8400B</u>		Serial Number: <u>73274</u>	<u>30</u> <u>May</u> <u>2024</u>
Initial Standby Value: <u>24+5984</u>		Ratio After Standardization: _____	
Cell Standardization (Daily)	Standardization Values:	Standard Info.	Bath Temperature: <u>24</u> °C
Conductivity must be ± 0.00001 compared to Standard	Before: _____	K ₁₅ VALUE: <u>0.999</u> <u>87</u>	Sample Temperature: <u>22</u> °C
	After: _____	Batch #: <u>P166</u>	Room Temperature: <u>21.5</u> °C
	Notes: _____	Batch Date: <u>6 Apr 2025</u>	Analyst: <u>CI</u>

[illegible]

Notes: used std info from 2024-079 run same day