

AUTOSAL ANALYSIS LOGSHEET

Cruise ID: <u>2024-028</u>	Cast/Station Name: <u>many</u>	Analysis Date (dd mmm yyyy): <u>23 Oct 2024</u>
Autosal Model: <u>3400B</u>	Serial Number: <u>73274</u>	
Initial Standby Value: <u>24 + 5987</u>	Ratio After Standardization: _____	
Cell Standardization (Daily)	Standardization Values:	Standard Info.
Conductivity must be ± 0.00001 compared to Standard 	Before: _____	K ₁₅ VALUE: 0.999 <u>87</u>
	After: _____	Batch #: <u>P166</u>
	Notes: _____	Batch Date: <u>10 Apr 2025</u>
		Bath Temperature: <u>24</u> °C
		Sample Temperature: <u>23</u> °C
		Room Temperature: <u>22</u> °C
		Analyst: <u>CI</u>

Sample	Station Name	Depth/Nisk#	Salinity 1	Salinity 2	Salinity 3	Comments
22	V107	0	1.84769	1.84771		
21	V107	5	1.84779	1.84778		
20	V107	10	1.84772	1.84773		
19	V107	25	1.84750	1.84750		
18	V107	50	1.86265	1.86262		
17	V107	75	1.88460	1.88459		
5003	100p 5003	100p	1.79745	1.77747		
16	V107	100-	1.91429	1.91429		
5002	100p 5002	m	1.84506	1.84504		
5001	100p 5001	3	1.82549	1.82548		
12	V107	400	1.95014	1.95019	1.95022	
11	V107	6-10	1.95177	1.95177		
14A	V107	200	1.94370	1.94368		cracked bottle lip
13	V107	300	1.94720	1.94720		
15	V107	150	1.93934	1.93936		
14B	V107	200	1.94376	1.94374		
54	BP01	5	1.84872	1.84872		
55	BP01	0	1.84861	1.84860		
52	BP01	25	1.86111	1.86109		
53	BP01	10	1.85005	1.85003		
50	BP01	75	1.91825	1.91826		
51	BP01	50	1.89606	1.89606		
49	BP01	100	1.93200	1.93200		
47	BP01	150	1.94250	1.94250		
48	BP01	125	1.94051	1.94049		Final Standby Value:
46	BP01	175	1.94413	1.94413		<u>24 + 5986</u>

Notes: <u>DWR</u>	read 1: <u>1.97871</u>	
	read 2: <u>1.97871</u>	

AUTOSAL ANALYSIS LOGSHEET

Cruise ID: <u>2024-028</u>		Cast/Station Name: <u>Nancy</u>	Analysis Date (dd mmm yyyy): <u>23 Oct 2024</u>
Autosal Model: <u>8400B</u>		Serial Number: <u>73274</u>	
Initial Standby Value: <u>24.5986</u>		Ratio After Standardization: _____	
Cell Standardization (Daily)	Standardization Values:	Standard Info.	Bath Temperature: <u>24</u> °C
Conductivity must be ± 0.00001	Before: _____	K ₁₅ VALUE: <u>0.999 87</u>	Sample Temperature: <u>23</u> °C
compared to Standard 	After: _____	Batch #: <u>P166</u>	Room Temperature: <u>22</u> °C
	Notes: _____	Batch Date: <u>6 Apr 2025</u>	Analyst: <u>CI</u>

[illegible]

Notes: _____

Notes: _____
